



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

LGC US Asset Holdings, LLC aka Lamons
7300 Airport Boulevard, Houston, TX 77061

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Mechanical Testing
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

Initial Accreditation Date:

December 20, 2020

Issue Date:

February 14, 2023

Expiration Date:

April 30, 2025

Accreditation No.:

111822

Certificate No.:

L23-117

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjllabs.com



Certificate of Accreditation: Supplement

LGC US Asset Holdings, LLC aka Lamons

7300 Airport Boulevard Houston TX 77061
Contact Name: Tommie Helms Phone: (713) 222-0284

Accreditation is granted to the facility to perform the following testing:

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Mechanical ^F	Metal Bolting and bolting materials	Hardness	ASTM A370 ASTM E18 (same process & method) Rockwell Hardness Tester & Hardness Reference Blocks	HRB Up to 100 HRC 20 to 60
	Metal Bolting and bolting materials	Tension Including tensile, yield, elongation, reduction of area	ASTM A370 ASTM E8 ASTM F606/F606M (same process and method) Tension test machine	Up to 400 000 lbf
	Metal Bolting and bolting materials	Charpy impact testing	ASTM A370 ASTM E23 (same process and method) Charpy impact tester (with cooling chamber)	-100 ° F to Ambient
	Metal Nuts	Proof load testing	ASTM A962 Tension test machine Both tension and Compression	Up to 400 000 lbf
	Nuts	Elevated temperature test	ASTM A194 Furnace to obtain specified temperature – Rockwell hardness tester	Up to 1 100 ° F Tested at ambient

1. The presence of a superscript F means that the laboratory performs testing of the indicated parameter at its fixed location. Example: Outside Micrometer^F would mean that the laboratory performs this testing at its fixed location.